

# MI CONSTANTE DE L.I.O.

	Dr. Aramberri	Dr. Mendicute	Dr. Soler	Dr. Zaldivar	Dr. Caballero	Dr. Cristóbal	Dr. Pascual	Dr. J.Fernández	Dr. R. Ruiz Mesa
<b>SN6AD1</b>	118.9	118.9	119	119.2	a0=0.523 a1=0.172 a2=0.211	118.6			118.9
<b>SN60TT</b>	118.6	118.6			a0=0.699 a1=0.119 a2=0.166	118.64			118.6
<b>SN60AT</b>	118.6	118.6	118.8	118.97	a0=0.09 a1=0.276 a2=0.163	118.9		118.7	118.6
<b>M Plus</b>	118.1	118.1				118.2			
<b>Tecnis MF</b>			119.1		a0=-1.065 a1=0.221 a2=0.232	119			119.1 (Acrílica)
<b>Rayner T Flex</b>						AL<25: 117.9 AL>25: 118.3			118
<b>Acri Lisa</b>				118		118.7		>25 mm: 118.30 <25 mm: 117.90	117.8
<b>AMO Ar40e</b>	118.6	118.6	118.7		a0=1.15 a1=0.40 a2=0.10	119	118.4		118.4
<b>Crystalent HD</b>						118.8		118.8	119
<b>ReZoom</b>			118.3		a0=0.92 a1=0.4 a2=0.1		118.3		118.4
<b>Biómetro</b>	IOL Master	IOL Master	IOL Master 5.4	IOL Master	IOL Master 5.4	IOL Master		IOL Master	IOL Master
<b>Queratómetro</b>	IOL Master	IOL Master	IOL Master TOPCON KR8100	IOL Master	Pentacam	IOL Master		IOL Master	IOL Master
<b>Fórmula</b>	SRK/T	SRK/T	SRK/T	SRK/T	HAIGIS	SRK/T		SRK/T (Ojos largos)  HOFFER Q (Ojos cortos)  HOLLADAY 2 (Lios premium)	HOLLADAY 2 HOFFER Q SRK/T